

**Statement on Prequalification of the Tender**  
**Tender No: 1D1/TSMSIDC/EQU-C/2016-17, Dated: 16-04-2016.**  
**Bids received on: 27.06.2016**

S.No	Tender Conditions	Information Furnished by the bidder	Remarks												
1	Firm Name and Address	<b>Page No: 1</b> M/s Biocare Medical Systems 16/11/20, Flat No: 401, Swathi Apartments, Saleem Nagar Colony, Malakpet, Hyderabad-500019. Phone No: 040-65608339,9949668339 Email Id: <a href="mailto:bcms99@yahoo.co.in">bcms99@yahoo.co.in</a>	Complies.												
2	Details of Quoted items :	<b>Page No : 56-57</b> Surgical Diathermy, Make: ALAN, Model: COMBIMAX- MBXXX	Complies.												
3	Manufacturer / Authorised Distributor. If authorized Distributor Name of the Manufacturer who given authorization	<b>Page No: 12</b> M/s Alan Electronic Systems Pvt Ltd Authorises M/s Bio Care Medical System.	Complies.												
4A	EMD required as per tender for quoted schedules Cumulative EMD required Item Name EMD Amount 1.Surigical Diathermy- 6000/- Total EMD- 6000/-	<b>Page No: 2</b> DD No:031359, dated: 31-05-16 for an amount of Rs.6,000/- from SBH	Complies.												
4B	Processing fee for Rs: 5725/-	<b>Page No: 3</b> DD No:031360, dated: 31-05-16 for an amount of Rs.5,725/- from SBH	Complies.												
5	General information on the bidder's company in <b>Form P-1</b>	<b>Page No: 1</b>	Complies.												
6	Total Sales of similar goods irrespective of brand and specification during each of Three financial years of, 2012-13 ,2013-4& 2014-15, ( <b>Form P2A</b> ): <b>Page No: 16</b>	<table border="1"> <thead> <tr> <th>Item</th> <th>Tender Qty</th> <th>PQ Qty</th> <th>2013-14</th> <th>2014-15</th> <th>2015-16</th> </tr> </thead> <tbody> <tr> <td>Surgical Diathermy</td> <td>1</td> <td>1</td> <td>70</td> <td>80</td> <td>90</td> </tr> </tbody> </table>	Item	Tender Qty	PQ Qty	2013-14	2014-15	2015-16	Surgical Diathermy	1	1	70	80	90	Complies.
Item	Tender Qty	PQ Qty	2013-14	2014-15	2015-16										
Surgical Diathermy	1	1	70	80	90										
7	Total sales of goods of model offered in Bid during each of Three financial years of 2012-13 ,2013-14 & 2014-15. <b>Form P2B Page No:17</b>	<table border="1"> <thead> <tr> <th>Item</th> <th>T. Qty</th> <th>PQ qty</th> <th>2013-14</th> <th>2014-15</th> <th>2015-16</th> </tr> </thead> <tbody> <tr> <td>Surgical Diathermy</td> <td>1</td> <td>1</td> <td>80</td> <td>90</td> <td>100</td> </tr> </tbody> </table>	Item	T. Qty	PQ qty	2013-14	2014-15	2015-16	Surgical Diathermy	1	1	80	90	100	Complies.
Item	T. Qty	PQ qty	2013-14	2014-15	2015-16										
Surgical Diathermy	1	1	80	90	100										
8	List of major customers in support of turnover in <b>Form P-3</b>	<b>Page No: 18-29</b> 1 Sri Sai Balaji Health Care India Pvt Ltd 2 Sravanthi Hospitals Rajahmundry 3 Dr. Bhavani Fertility Centre	Not Complies.  Supporting Documents enclosed not for the quoted Model												
9	Details of service centres in TS in <b>Form P-4</b>	<b>Page No:30</b> One service Centre with 5 Support Engineers.	Complies.												
10	Past supply Performance certificate in <b>Form-P5</b> (Certificate from statutory auditor)	<b>Page No : 31</b>	Complies.												
11	Financial Capacity of Bidder in <b>Form P6, Page No: 32-36</b>	<table border="1"> <thead> <tr> <th>2013-14</th> <th>2014-15</th> <th>2015-16</th> <th>Avg. Annual Turnover</th> </tr> </thead> <tbody> <tr> <td>5 Lakhs</td> <td>1.72Cr</td> <td>2.10Cr</td> <td>2.20Cr</td> </tr> </tbody> </table>	2013-14	2014-15	2015-16	Avg. Annual Turnover	5 Lakhs	1.72Cr	2.10Cr	2.20Cr	Complies.  Balance Sheets Enclosed.				
2013-14	2014-15	2015-16	Avg. Annual Turnover												
5 Lakhs	1.72Cr	2.10Cr	2.20Cr												

S.No	Tender Conditions	Information Furnished by the bidder	Remarks
12	Declaration in Form 7	Page No:37	Complies.
13	Form T1 Specifications (Quoted technical Higher/Lower)	Page No:38-40 Specifications given for High end with Vessel ceiling machine. Boucher enclosed against other than quoted model.(Model: Elsy-360 M+)	Not Complies.
14	Form T2 (Check List)	Page No: 41-42	Complies.
15	Form T3 (List of Documents Scanned and Uploaded)	Page No: 55	Complies.
16	Form-T4 (Documents to be submitted in hard copy in MSIDC office)	Page No:58	Complies.
17	Pre-qualification criterion	Page No:44-47	Complies.
18	Technical specifications	Page No: 46-47	Complies.
19	General instruction to bidders	Page No:48	Complies.
20	Standard procedure for bid evaluation	Page No:49	Complies.
21	General condition of proposed contract(GCC)	Page No:50-51	Complies.
22	Special Conditions	Page No:52	Complies.
23	Delivery Period Within 60 days	Page No:43	Complies.
24	Bid Letter Form	Page No:53	Complies.
25	Section-C	Page No: 54	Complies
26	The equipment should be USFDA and European-CE certified	Page No:3-5, 9. ISO certificates and CE enclosed. <b>USFDA not furnished.</b>	Not Complies.
27	Any legal Cases or Blacklisted	No adverse complaints.	Complies.
28	Other Documents	Page No: 10 & 11 VAT registration with Telangana State TIN No: 36229584951 PAN No:AAIFB7556N	Complies

**Remarks:**

1. Details submitted in respect to P3 form are other than quoted model.
2. P3 not satisfies P2B
3. Satisfactory performance certificates enclosed for Model Elsy-360 against quoted model Combimax –MBXXX.
4. Only CE Certificate furnished against asked CE & US FDA.